



CHYMOTRYPSIN

DETERMINATION OF CHYMOTRYPSIN IN FAECES

Colorimetric method

Suitable for all analyzers – 72 tests


Product insert with instructions for automated and manual procedures

2x level Chymotrypsin controls



SUMMARY

PRINCIPLE



SAMPLE MATERIAL

Faeces

SAMPLE PREPARATION

Use sample preparation tubes (product nr.: 2505) or centrifugation tubes with caps.

LINEARITY

Up to 60 U/gram faeces

EXPECTED VALUES IN FAECES

	30 °C	37 °C
Normal	> 8.4 U/g	> 13.2 U/g
Further investigation necessary	4.2 – 8.4 U/g	6.6 – 13.2 U/g
Pathologic	< 4.2 U/g	< 6.6 U/g

QUALITY CONTROL

Products	Product no.	Quantity
Chymotrypsin Control Low Level	2910	10 x 2 ml
Chymotrypsin Control High Level	2986	10 x 2 ml

QUANTITY OF DETERMINATIONS

According to product insert of Chymotrypsin Reagent Set (2909):

Procedure

- Manual : 18 tests
- Automated : 72 tests

PRODUCTS

Products	Product no.	Quantity
Chymotrypsin Reagent Set	2909	18 manual 72 automated
Chymotrypsin Control Low Level	2910	10 x 2 ml
Chymotrypsin Control High Level	2986	10 x 2 ml
Chymotrypsin Solvent Solution	2912	1 x 500 ml
Faecal Sample Preparation Set Tubes	2505	50 tests



NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstruChemie for the complete product insert and latest edition.

Printed in the Netherlands, August 2018, Edition 4.0