



FREE GLYCEROL

DETERMINATION OF FREE GLYCEROL IN SERUM, PLASMA AND FOODSTUFFS

Colorimetric method

Suitable for all analyzers – 200 tests




Product insert with instructions for automated and manual procedures

Incl. Glycerol Standard

For automated analyzers.

SUMMARY

PRINCIPLE



SAMPLE MATERIAL

Serum, heparinized or EDTA-plasma

LINEARITY

Up to 11.40 mmol/l Glycerol

EXPECTED VALUES

Serum, plasma: 0.03 – 0.19 mmol/l Glycerol

QUANTITY OF DETERMINATIONS

According to product insert of Glycerol Reagent Set (2913):

Procedure

- Manual : 50 tests
- Automated : 200 tests

PRODUCTS

Products	Product no.	Quantity
Glycerol Reagent Set	2913	50 manual 200 automated



NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstru**chemie** for the complete product insert and latest edition.

Printed in the Netherlands, August 2018, Edition 4.0