



GALACTOSE

DETERMINATION OF GALACTOSE IN WHOLE BLOOD OR URINE

Enzymatic method

Suitable for all analyzers – 450 tests

Product insert with instructions for automated and manual procedures


Stability > 6 years after production

Galactose controls available

With Galactose intake tolerance test

SUMMARY

PRINCIPLE



SAMPLE MATERIAL

Whole Blood or Urine

LINEARITY

Up to 4 mmol/l

EXPECTED VALUES

Expected Values in fasting whole blood: Up to 240 µmol/l Galactose.

Expected Values in Urine of neonates: < 3.3 mmol (24 hours urine) Galactose.

Expected Values in Urine of older children and/or adults: < 0.2 mmol/l Galactose.

QUALITY CONTROL

Products

Galactose Control Normal Level

Galactose Control High Level

Product no.

2939

3022

Quantity

1 x 5 ml

1 x 5 ml

QUANTITY OF DETERMINATIONS

According to product insert of Galactose Reagent Set (2935):

Procedure

- Manual : 30 tests

- Automated : 450 tests

PRODUCTS

Products	Product no.	Quantity
Galactose Reagent Set	2935	35 manual 450 automated
Galactose Calibrator	3021	1 x 5 ml
Galactose Control Normal Level	2939	1 x 5 ml
Galactose Control High Level	3022	1 x 5 ml
Perchloric Acid 0.33 mol/l Deproteinization Reagent	2934	1 x 100 ml



NOTES

1. For in vitro diagnostic use only.
2. For professional use only.
3. Always contact INstru**chemie** for the complete product insert and latest edition.

Printed in the Netherlands, August 2018, Edition 4.0